

## Corrigenda

Please note the following corrigenda on tender Ref no STIC/SAIF/MALDI/2018-19/2092

SI no	Section /Clause No	Item no	Page no	Clause/Item in Tender	Amend to read as /new addition
1	Section V	1.Mass analyzer-Linear mode	40	Mass Accuracy Better than 2 PPM	Deleted as not applicable in linear mode. Added new term as Mass Accuracy better than $\leq 30$ PPM with internal calibration and $\leq 200$ ppm with external calibration
2	Section V	1.Mass analyzer-Linear mode	40	Measureable mass range in this mode m/z 1-300kDa	Measureable mass range in this mode m/z 1-500kDa
3.	Section V	1.Mass analyzer-reflection mode	41	Mass Accuracy Better than $\leq 2$ PPM with internal calibration and $\leq 10$ ppm with external calibration	Mass Accuracy Better than $\leq 10$ PPM with internal calibration and $\leq 100$ ppm with external calibration
4	Cover page		2	Brief description of stores /goods/equipment  Bench top MALDI TOF and accessories	Read as MALDI TOF mass spectrometer and accessories. Deleted the term Bench top